

Doc. No.	KSX/TD-MEBN-017	Title	EU Declaration of Conformity of Nonsterile Emergency Blanket		
Ver./Rev.No.	A/0	Issued Date	2024.08.30	Page/Total	1 / 3

EU Declaration of Conformity

Manufacturer Name: Kingstar Medical (Xianning) Co., Ltd.

Manufacturer Address: No. 79 Yong'andong Road, Xian'an District 437100, Xianning City, Hubei Province, the People's Republic of China

SRN of the Manufacturer: CN-MF-000006015

Location of Manufacturer: Xianning City, Hubei Province, China.

Authorized Representative: Shanghai International Holding Corp. GmbH (Europe)

SRN of the Authorized Representative: DE-AR-000000001

Address of their Registered Place of Business: Eiffestraße 80, 20537 Hamburg, Germany

Location be established: Germany

Basic UDI-DI: 6971872201491000PY

Name of the device: Nonsterile Emergency Blanket

EMDN Code: T030199, Covers, Instruments and Equipment - Other

UMDN Code: 10416, Blankets, Aluminized Rescue

GMDN Code: 66298, Rescue blanket, single-use

Intended Purpose: It is intended to keep a person warm and to prevent the further loss of body heat in an emergency rescue situation. It is single-use device.

Risk Class of the Device: Class I based on Rule 1 of ANNEX VIII of Regulation (EU) 2017/745.

All non-invasive devices are classified as class I.

The conformity assessment procedure performed: According to Article 19 of the Regulation (EU) 2017/745, draw up this EU declaration of conformity which contain the information set out in Annex II and III of the Regulation (EU) 2017/745.

CS used or Standard applied: Please find in Annex II.

Identification of the device: Please find in Annex I.

Declaration: This declaration of conformity is issued under the sole responsibility of Kingstar Medical (Xianning) Co., Ltd. We hereby declare that the medical device specified above meet the provision of the Regulation (EU) 2017/745 for medical device. This declaration is supported by the quality system approval to ISO 13485 by TÜV SÜD Product Service GmbH.

All supporting documentation is retained at the premises of the manufacturer.

Notified Body: TÜV SÜD Product Service GmbH

Address: Ridlerstr. 65, 80339 Munich, Germany

Identification No.: CE0123


Signed for and on behalf of:

Place of Issue: Xianning City, Hubei Province, China.

Date of Issue: 2024-08-30

Print Name: Fan Rong

Function: Management Representative

Signature: 

Doc. No.	KSX/TD-MEBN-017	Title	EU Declaration of Conformity of Nonsterile Emergency Blanket		
Ver./Rev.No.	A/0	Issued Date	2024.08.30	Page/Total	2 / 3

Annex I --- Identification of the Device Covered by the EU Declaration of Conformity

1. Identification of the Device

Table --- Identification of the Device

No.	WERO REF	Kingstar REF	Kingstar Art. Nr.	Product name	Specification	Packaging configuration
1	210120	C040004	5142420040	Nonsterile Emergency Blanket	160cm×210cm, Gold/silver	1 pc/PE bag, 250 bags/carton
2	210186	C040005	5142420020		160cm×210cm, Green/silver	1 pc/foil bag, 300 bags/carton

2. Photograph of Nonsterile Emergency Blanket



Photo 1 --- Nonsterile Emergency Blanket



Photo 2 --- Nonsterile Emergency Blanket



Photo 3 --- Emergency Blanket in nonsterile packaging

Annex II --- European Harmonization and International Standard list

Category	No.	Standards	Content
QMS	1	EN ISO 13485:2016/A11:2021	Quality management systems — Requirements for regulatory purposes (ISO 13485:2016)
Labeling	2	ISO 15223-1: 2021	Medical devices - Symbols to be used with information to be supplied by the manufacturer- Part1: General requirements
	3	EN ISO 20417: 2021	Medical devices- Information to be supplied by the manufacturer (ISO 20417:2021)
Risk management	4	EN ISO 14971:2019/A11:2021	Medical devices-Application of risk management to medical devices (ISO 14971:2019)
Usability	5	IEC 62366-1:2015/A1:2020	Medical devices-Application of usability engineering to medical devices
Clinical Evaluation	6	MEDDEV 2.7.1 rev 4	Guidance document for clinical evaluation
Sampling Inspection	7	ISO 2859-1:1999/A1:2011	Sampling procedures for inspection by attributes--Part 1: Sampling schemes indexed by acceptance quality limit (AQL) for lot-by-lot inspection
Transportation	8	ISTA 2A: 2011	Partial simulation performance test procedure

Doc. No.	KSX/TD-MEBN-017	Title	EU Declaration of Conformity of Nonsterile Emergency Blanket		
Ver./Rev.No.	A/0	Issued Date	2024.08.30	Page/Total	3 / 3

PMS	9	MEDDEV 2.12-1:2013	Guidelines on a medical devices vigilance system
PSUR	10	ISO/TR 20416: 2020	Medical devices - Post-market surveillance for manufacturers
Environment	11	EN ISO 14644-1: 2015	Cleanrooms and associated controlled environments — Part 1: Classification of air cleanliness by particle concentration (ISO 14644-1:2015)
	12	EN ISO 14644-2: 2015	Cleanrooms and associated controlled environments — Part 2: Monitoring to provide evidence of cleanroom performance related to air cleanliness by particle concentration (ISO 14644-2:2015)
	13	EN ISO 14644-3: 2019	Cleanrooms and associated controlled environments — Part 3: Test methods (ISO 14644-3:2019)
Biocompatibility	14	EN ISO 10993-1: 2020	Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process (ISO 10993-1:2018)
	15	EN ISO 10993-5: 2009	Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity (ISO 10993-5:2009)
	16	EN ISO 10993-10:2023	Biological evaluation of medical devices — Part 10: Tests for skin sensitization (ISO 10993-10:2021)
	17	EN ISO 10993-11: 2018	Biological evaluation of medical devices — Part 11: Tests for systemic toxicity (ISO 10993-11:2017)
	18	ISO 10993-18:2020/A1:2022	Biological evaluation of medical devices — Part 18: Chemical characterization of medical device materials within a risk management process
	19	EN ISO 10993-23:2021	Biological evaluation of medical devices - Part 23: Tests for irritation (ISO 10993-23:2021)