



DECLARATION OF CONFORMITY

ACCORDING TO (EU) 2017/745 MEDICAL DEVICE REGULATION

EU Representative

SUNGO Europe B.V.
Fascinatio Boulevard 522, Unit 1.7,
2909VA Capelle aan den IJssel, The
Netherlands
SRN: NL-AR-000000247

Conformity Assessment

Conformity Assessment Procedure
Annex II+III of Regulation (EU) 2017/745

Applicable Standards
EN ISO 14971: 2019
EN ISO 15223-1: 2021
EN ISO 20417: 2021
EN ISO 10993-1: 2020
EN ISO 10993-5: 2009
EN ISO 10993-10: 2023
EN ISO 10993-23: 2021

Remark

*The declaration of conformity is valid in connection
with the release technical document
CE/MDR-Q030199-11.*

*All the supporting documentation is retained at the
premises of the manufacturer.*

*The Declaration of Conformity is exclusively under
the sole responsibility of the manufacturer.*

Manufacturer

Name: ZHANGJIAGANG XIEHE MEDICAL
APPARATUS&INSTRUMENTS CO.,LTD.
Address: No.7th, Middle Xinzha Road, Zhashang
Industrial Zone, Yangshe Town, Zhangjiagang City,
Jiangsu 215600, China
SRN: CN-MF-000008449

Product Information

Name: Manual Suction
Model: YXH-MSR, YXH-M01, XH-KB04, XH-KB05,
XH-KB06, XH-KB07, XHKB09, XH-KB10
EMDN: Q030199
Basic UDI-DI: 6974580870007E2
Classification: Class I, According to Rule 5, Annex
VIII, Regulation (EU) 2017/745

Declaration

We herewith declare that the above-mentioned
products meet the requirements of Medical Device
Regulation (EU) 2017/745 and the applicable
standards above.

Signature:

Date: 2024.11.22

Position: GM

Place: Zhangjiagang/China

