

## EU Declaration of conformity no. 210422-006

Product Name:	Vacuum Hand Pump FW
Intended use:	Intended use for evacuation of Germa Vacuum Splints and Germa Positioning Pillows for immobilization and positioning of patients.
SRN:	SE-MF-000003932
Basic UDI-DI:	735001959P01ACCESSO48
UDI DI:	07350019595497
Germa Article No:	23272000701
Manufacturer:	AB Germa
Visiting address:	Industrigatan 54-56, SE-29136 Kristianstad
Phone:	+46 (0)44 123030
Email:	info@germa.se
Web:	www.germa.se
Product class:	Class I according to rule 1 in Annex VIII in MDR 2017/745
Conformity procedure:	Self-certification according to Annex IV in MDR 2017/745
Identification:	All products with serial numbers issued from;
	LOT number: 517470
	Date: 2021-05-25 (yyyy-mm-dd)

## Declaration statement;

This EU declaration of conformity is issued under the sole responsibility of the manufacturer AB Germa. The devices covered by this declaration is in conformity with the requirements in the European Medical Device Regulation 2017/745.

Signed in Kristianstad, Sweden on behalf of AB Germa by:

Position: Managing Director Name: Björn Holmqvist

Date: 2021-05-23

Postadress/Postal address

Adress/Goods address

E-mail/Web page

VAT No.

AB GERMA Box 2127 SE-291 02 KRISTIANSTAD SWEDEN

Industrigatan 54-56 SE-291 36 Kristlanstad +46 (0)44-123030

+46 (0)44-10 31 79

info@germa.se www.germa.se SE556086941301

481-4125