

Declaration of Conformity

Product Name	Saver Safe Evacuation Chair	Ref No:	DOC – SAV
Product Code/GTIN No	SAV-001	Issue No:	1
Intended Use	Patient Transportation/Evacuation	Review Date:	Nov 2020
Product Classification	Class I Medical Device as per Rule 1 Chapter III Annex VIII MDR 2017/745		
Manufacturers Name	Ferno (UK) Limited		


Manufacturers Address & UKCA Representative		EU Authorised Representative (if applicable)	
Ferno House, Stubs Beck Lane, Cleckheaton, West Yorkshire. BD19 4TZ United Kingdom Tel: +44 (0) 1274 851999 www.ferno.co.uk		Ferno-Washington Italia s.r.l. Via B. Zallone, 26, Pieve di Cento 40066 Bologna Italy Tel: +39.051.6860028 www.ferno.it	

LOT ID/UDI prefix SAV ##### **Date of First Manufacture** 2009


Serial Numbers are identified on the Manufacturing and Inspection Record and/or the Customer Database. These should be referenced in all communication as customer variants, including SM & PN, affect the Product Code reference.

The undersigned hereby declare that the medical device specified above complies with the General Safety and Performance Requirements within *Annex I* of EC Regulation 2017/745, BS EN 1789 Road Ambulances and their equipment and additional designated standards referenced to compile the technical documentation detailed in the 'Technical File' as held by the Engineering & Design Manager at the above address.

This declaration is compiled in accordance with *Article 19* and *Annex IV* of EC Regulation 2017/745 and supported by a Quality Management System which has been assessed and approved to BS EN ISO 9001:2015 by ACM-CCAS Quality Assurance assessment services.

 **01 December 2020** **Shahid Saleem**

Signed: Engineering & Design Manager Dated Print Name

 **01 December 2020** **Jon Ellis**

Signed: Managing Director Dated Print Name